

VNMR Parameters

Query / Setting: sw? sw=3000 gain='n'

Spectral Window:

sw	sfrq	tof
spectral width	spectr. frequency	transmitter off set

FID acquisition:

np	at	gain
number of points	acquisition time	receiver gain

Decoupling:

dof	dm	dpwr
decoupler off set	decoupler mode	decoupler power

Processing:

lb	fn
line broadening	Fourier number

Flags:

dp	il	(y or n)
double precision	interleave	

Special:

temp
sample temperature